







LOCKING IMPLANTS SYSTEM - HAND

- *SURGICAL TECHNIQUE*
- *IMPLANTS*
- *INSTRUMENT SET*



SYMBOLS DESCRIPTION

	Titanium or titanium alloy		H length [mm]
	Cobalt		Angle
	Left		available lengths
	Right		Available number of holes
	Available versions: left/right		Thickness [mm]
	Length		Scale 1:1
	Torx drive		Number of threaded holes in the shaft part of the plate
	Torx drive cannulated		Number of locking holes in the plate
	Hexagonal drive		Variable angle
	Hexagonal drive cannulated		Cortical
	Cannulated		Cancellous
	Locking		Available in sterile/ non- sterile condition
	Diameter [mm]		Refer to surgical technique



Caution - pay attention to a special procedure.



Perform the activity under X-Ray control.



Information about the next stages of a procedure.



Proceed to the next stage.



Return to the specified stage and repeat the activity.



Before using the product, carefully read the Instructions for Use. It contains, among others, indications, contraindications, side effects, recommendations and warnings related to the use of the product.



The above description is not a detailed instruction of conduct. The surgeon decides about choosing the operating procedure.

www.chm.eu

Document No ST/80-301
Date of issue 14.03.2022
Review date P-001-23.01.2023

The manufacturer reserves the right to introduce design changes.
Updated INSTRUCTIONS FOR USE are available at the following website: ifu.chm.eu

1. INTRODUCTION	5
2. IMPLANT DESCRIPTION	6
3. INSTRUMENTS DESCRIPTION	8
4. SURGICAL TECHNIQUE	9
4.1. FRACTURE REDUCTION	9
4.2. IMPLANT SELECTION	9
4.3. TEMPORARY PLATE STABILIZATION	9
4.4. SCREWS INSERTION IN THE EPIPHYSEAL PART OF THE PLATE	9
4.5. SCREWS INSERTION IN THE OVAL-SHAPED HOLE OF THE SHAFT PART OF THE PLATE	9
4.6. SCREWS INSERTION IN THE SHAFT PART OF THE PLATE	9
4.7. WOUND CLOSURE	9
5. SURGICAL PROCEDURES	10
5.a. PROCEDURE OF A SCREW INSERTION INTO THE LOCKING HOLE	10
5.b. PROCEDURE OF SCREW INSERTION INTO THE OVAL-SHAPED HOLE	12
5.c. PLATE SHAPING	14
5.d. PLATE TRIMMING	15
6. LAG SCREW SURGICAL TECHNIQUE	16
6.1. IMPLANT SELECTION	16
6.2. GLIDING HOLE DRILLING	16
6.3. CORE HOLE DRILLING	17
6.4. HOLE DEEPENING FOR THE SCREW'S HEAD	17
6.5. HOLE DEPTH MEASURE	18
6.6. SCREW INSERTION	18
6.7. WOUND CLOSURE	19
7. POSTOPERATIVE PROCEDURE	19
8. IMPLANT REMOVAL	19
9. CATALOGUE PAGES	20
9.a. INSTRUMENT SET	20
9.b. STAND FOR PLATES	24
9.c. STAND FOR SCREWS	27
9.d. PLATES	29
9.e. SCREWS	32

1. INTRODUCTION

This surgical technique applies to 3.0ChLP locked plating system used for stabilization of small bones of the hand. The plates are a part of the ChLP locked plating system developed by **ChM**. The presented range of implants is made of materials in accordance with ISO 5832 standards.

The system includes:

- implants (*plates and screws*),
- instrument set used in the surgery,
- surgical technique.

Indications

- fractures of small bones of the hand,
- bone fracture mal-, and non-unions,
- surgical corrections, osteotomies.

Plate selection and shaping

The plates are available in various lengths and widths, for left and right limb separately. This allows for optimal selection of the implant to the fracture type. Shaping of the plates is allowed.



Before using the product, carefully read the Instructions for Use. It contains, among others, indications, contraindications, side effects, recommendations and warnings related to the use of the product.



The above description is not a detailed instruction of conduct. The surgeon decides about choosing the operating procedure.

2. IMPLANT DESCRIPTION

Small plates are a part of 3.0ChLP system. This system includes also compatible locking screws. To facilitate their identification, both titanium plate and screws are marked grey.

0.8




1.2



Anatomical design:

- adapted to the bone structure,
- 2 options of the plate thickness and width for optimal match,
- soft top edges significantly reduce irritation of peri-implant tissues.




Locking holes:

- for use with locking screws,
- 15° locking screws angulation from the predefined axis of the hole,
- also for use with non-locking screws.


Kirschner wire holes:

- for secure temporary plate stabilization,
- for easier plate positioning.










Oval-shaped hole:

- for fracture compression,
- for use with locking screws,
- also for use with non-locking screws.



Features of 3,0ChLP screws system:


- optimal implant selection - all system screws are compatible with all holes in the plate,
- easy identification - screws marked as specified in the table (Tab.1) below.
- color coding - a color has been assigned to each screw diameter and compatible instrument used for its implantation.

Screw diameter	Screw color	VA locking screws	Non-locking screws
1.2mm	Blue	—	
1.5mm	Yellow		
2.0mm	Green		
2.3mm	Brown		


Tab. 1. Table of colors assigned to the screws

All 3,0ChLP screws are packed in collective packages (*clip magazines*) - 5 screws of a given length in each clip. The clip is also provided with additional information to facilitate the identification of the screws:

- type of screws (*VA LOCKING, CORTICAL*),
- screws diameter (1.2, 1.5, 2.0, 2.3)
- screws length,
- catalog and LOT number.



Clip magazines with screws are distributed on the stand for 3,0ChLP screws 4x2H 1/2H [14.0203.501].



3. INSTRUMENTS DESCRIPTION

The instruments are marked with the color assigned to the screw diameter. Drills and guides have 1 or 2 color stripes:


- single color stripe - the instrument for drilling the core holes,
- double color stripes - the instrument for drilling the gliding hole.

Color coded instruments list:

Screw diameter	Screw color	Core hole drill (Core hole)	Gliding hole drill (Gliding hole)	Double drill guide	Core hole drill guide (Core hole)	Gliding hole drill guide (Gliding hole)
1.2mm	Blue					
1.5mm	Yellow					
2.0mm	Green					
2.3mm	Brown					

The stand for screws has been equipped with:

- holes for screws diameter verification,
- color markers to facilitate the selection of compatible instruments,
- a measure for screws length verification.



4. SURGICAL TECHNIQUE

4.1. FRACTURE REDUCTION





Perform fracture reduction. If need be, temporarily stabilize the bone fragments with Kirschner wires, bone holding forceps and/or reduction pliers.

4.2. IMPLANT SELECTION

Select the right implant to the type and place of fracture, bone size and structure. If necessary, shape (*acc. to 5c point*) and trim (*acc. to 5d point*) the plate

4.3. TEMPORARY PLATE STABILIZATION

Stabilize the position of the implant inserting guide rods 0.8/150 [40.8132.150] into dedicated holes of the plate.



Perform X-Ray control of the fracture reduction

4.4. SCREWS INSERTION IN THE EPIPHYSEAL PART OF THE PLATE

Insert locking or cortical screw of a suitable length and diameter into the locking holes of the epiphyseal part of the plate (*acc. to procedure 5b*).

4.5. SCREWS INSERTION IN THE OVAL-SHAPED HOLE OF THE SHAFT PART OF THE PLATE

Insert locking or cortical screw of a suitable length and diameter into the oval-shaped hole of the shaft part of the plate (*acc. to procedure 5a*). Compress the fracture fragments if necessary.

4.6. SCREWS INSERTION IN THE SHAFT PART OF THE PLATE

Insert locking or cortical screws of a suitable length and diameter into the locking holes of the shaft part of the plate (*acc. to procedure 5b*).



Insert the cortical screws into the fracture before inserting the locking screws.

The doctor decides about the order and number of locking and cortical screws to be inserted.

4.7. WOUND CLOSURE

Before closing the wound, take an X-Ray image in at least two projections to confirm implant position and fracture reduction. Make sure all the screws are properly tightened and do not penetrate the joint surface. Use appropriate surgical technique to close the wound.

5. SURGICAL PROCEDURES

5.a. PROCEDURE OF A SCREW INSERTION INTO THE LOCKING HOLE


Position the drill guide

Position the drill guide in the hole.



Use instruments marked according to the table in section 3. INSTRUMENTS DESCRIPTION.

Exceeding the 15° angulation may prevent the correct locking of the VA screw in the hole.



	40.8463.000	1.2
	40.8464.000	1.5
	40.8465.000	2.0
	40.8466.000	2.3

Hole drilling


Use a dedicated drill to drill a hole in a desired position and determined depth.



Use instruments marked according to the table in section 3. INSTRUMENTS DESCRIPTION.



Drill under X-Ray control, to avoid the collision of the drill with the already implanted screws.




	40.8467.660	1.2
	40.8468.670	1.5
	40.8470.670	2.0
	40.8479.670	2.3

Measurement of hole depth

Insert depth measure [40.4640.500] into drilled hole until the hook of the measure rests against the outer surface of the other cortex.

	40.4640.500
--	-------------







Screw insertion

Use quick coupling handle [40.6405.300] and screwdriver tip X [40.6449.057] or screwdriver tip X with holder [40.8475.000] to insert the screw.



If screwdriver tip X with holder [40.8475.000] is used, remember to remove the holder from the screw head for final locking it in the plate.

	40.6449.057
	40.8475.000
	40.6405.300



5.b. PROCEDURE OF SCREW INSERTION INTO THE OVAL-SHAPED HOLE



Position the drill guide

Position the drill guide in the desired position (*compression or neutral*).



Use instruments marked according to the table in section 3. INSTRUMENTS DESCRIPTION.

	40.8463.000	1.2
	40.8464.000	1.5
	40.8465.000	2.0
	40.8466.000	2.3




Hole drilling

Use a drill dedicated to the selected screw diameter and positioned as desired to drill a hole.



Use instruments marked according to the table in section 3. INSTRUMENTS DESCRIPTION.

	40.8467.660	1.2
	40.8468.670	1.5
	40.8470.670	2.0
	40.8479.670	2.3




Measurement of hole depth

Insert depth measure **[40.4640.500]** into drilled hole until the hook of the measure rests against the outer surface of the other cortex.



40.4640.500



Screw insertion

Shaping of the plates should only be performed with the use of plate bending pliers **[40.8460.100]**.



If screwdriver tip X with holder **[40.8475.000]** is used, remember to remove the holder from the screw head for final locking it in the plate.



40.6449.057



40.8475.000



40.6405.300



5.c. PLATE SHAPING


Shaping of the plates should only be performed with the use of plate bending pliers **[40.8460.100]**.



40.8460.100

Place the two adjacent holes of the plate in the jaws and tighten the pliers.

Bend the plate as required, but do not exceed $20^\circ \div 25^\circ$ at a time. If necessary, repeat shaping by positioning the plate on the next two holes.



5.d. PLATE TRIMMING



For trimming, only the plate cutting pliers **[40.8406.100]** can be used.



40.8406.100

Determine the required implant length and position the plate hole on the positioning pin.

Close the arms of the pliers to cut the plate. When cutting, hold the cut-off part of the plate. The trimmed plate will remain on the positioning pin.



Sharp edges removal

Rotate the plate by 180°. Position the last hole of the plate on the positioning pin. Close the arms of the pliers again. Sharp edges will be dulled.

6. LAG SCREW SURGICAL TECHNIQUE

6.1. IMPLANT SELECTION

Select the appropriate screw diameter for the type and location of the fracture, bone size and structure.

6.2. GLIDING HOLE DRILLING

Drill guide installation

Screw in the dedicated drill guide into the sleeve of the reduction forceps **[40.8461.000]**.



Use instruments marked according to the table in section 3. INSTRUMENTS DESCRIPTION.



40.8461.000



40.8462.120

1.2



40.8462.150

1.5



40.8462.200

2.0



40.8462.230

2.3

Forceps positioning on a bone

Set the reduction forceps on the bone in a position appropriate for the type and location of the fracture. Close the arms of the forceps with adequate force.

Hole drilling

Using a drill dedicated to the selected screw, make a gliding hole to the desired depth.



Use instruments marked according to the table in section 3. INSTRUMENTS DESCRIPTION.



40.8468.660

1.2



40.8469.670

1.5





40.8471.670

2.0



40.8472.670

2.3







6.3. CORE HOLE DRILLING

Drill guide installation

Screw in the dedicated drill guide into the sleeve of the reduction forceps **[40.8461.000]**.



Use instruments marked according to the table in section 3. INSTRUMENTS DESCRIPTION.

	40.8462.080	1.2
	40.8462.120	1.5
	40.8462.150	2.0
	40.8462.180	2.3

Hole drilling





Using a drill dedicated to the selected screw, make a core hole to the desired depth.



Use instruments marked according to the table in section 3. INSTRUMENTS DESCRIPTION.






Drill under X-Ray control, to avoid the collision of the drill with the already implanted screws.

	40.8467.660	1.2
	40.8468.670	1.5
	40.8470.670	2.0
	40.8479.670	2.3

6.4. HOLE DEEPENING FOR THE SCREW'S HEAD


Remove the drill guide. Use quick coupling handle **[40.6405.300]** and counter-sink **[40.8478.000]** to deepen the hole for the screw's head.

	40.8478.000
	40.6405.300



6.5. HOLE DEPTH MEASURE

Insert the depth measure **[40.4640.500]** into the drilled hole until the hook of the measure reaches the outer surface of the other cortex.





6.6. SCREW INSERTION

Use quick coupling handle **[40.6405.300]** and screwdriver tip X **[40.6449.057]** or screwdriver tip X with holder **[40.8475.000]** to implant the screw.



If screwdriver tip X with holder **[40.8475.000]** is used, remember to remove the holder from the screw head for final locking it in the plate.



6.7. WOUND CLOSURE

Before closing the wound, take an X-Ray image in at least two projections to confirm implants position and fracture reduction. Make sure all the screws are properly tightened and do not penetrate the joint surface. Use appropriate surgical technique to close the wound.

7. POSTOPERATIVE PROCEDURE

Introduce appropriate postoperative treatment. The physician decides on the post-operative treatment and its conduct. In order to avoid patient's movement limitations, introduce exercises as soon after surgery as possible. However, make sure that the limb is not fully loaded before fragments osteosynthesis is complete.

8. IMPLANT REMOVAL

The physician decides about implant removal. In order to remove the implants from the body, use quick coupling handle **[40.6405.300]** and screwdriver tip X **[40.6449.057]** to unlock all the locking screws first and then remove them from the bone. This will prevent any rotation of the plate when removing the last locking screw.

	40.6449.057
	40.6405.300

9. CATALOGUE PAGES

9.a. INSTRUMENT SET













Set for 3.0ChLP 4x4 H

15.0203.001

	Name	Catalogue No.	Pcs
	Instrument set for 3.0ChLP 4x4 1/2H	15.0203.201	1
	Instrument set for 3.0ChLP 4x2 1/2H	15.0203.202	1
	Instrument set for 3.0ChLP 4x2 1/2H	15.0203.203	1
	3.0ChLP container lid 4x4H	14.0203.102	1
	Container 3.0ChLP 4x4H	14.0203.101	1



















Instrument set for 3.0ChLP 4x4 1/2H

15.0203.201

	Name	Catalogue No.	Pcs
	Tray for 3.0ChLP instrument set 4x4 1/2H	14.0203.201	1
	Bending forceps for plates	40.8460.100	2
	Bone holding forceps 135mm	40.4150.000	1
	Bone holding forceps 90mm	40.4150.090	1
	Plate cutting pliers	40.8406.100	1
	Reduction forceps	40.8461.000	1
	Drill guide 0.9	40.8462.090	1
	Drill guide 1.2	40.8462.120	1
	Drill guide 1.5	40.8462.150	1
	Drill guide 2.0	40.8462.200	1
	Drill guide 1.8	40.8462.180	1
	Drill guide 2.3	40.8462.230	1









Instrument set for 3.0ChLP 4x2 1/2H

15.0203.202

	Name	Catalogue No.	Pcs
	Stand for instrument set for 3.0ChLP 4x2 1/2H	14.0203.202	1
	Drill guide 09/1.2	40.8463.000	1
	Drill guide 1.2/1.5	40.8464.000	1
	Drill guide 1.5/2.0	40.8465.000	1
	Drill guide 1.8/2.3	40.8466.000	1
	Drill 0.9/60	40.8467.660	1
	Drill 1.2/60	40.8468.660	1
	Drill 1.2/70	40.8468.670	1
	Drill 1.5/70	40.8469.670	1
	Drill 1.5/70	40.8470.670	1
	Drill 2.0/70	40.8471.670	1
	Drill 1.8/70	40.8479.670	1
	Drill 2.3/70	40.8472.670	1
	Depth measure	40.4640.500	1
	Countersink	40.8478.000	1
	Screwdriver tip X	40.6449.057	1
	Screwdriver tip X with holder	40.8475.000	1
	Quick coupling handle	40.6405.300	1

Instrument set for 3.0ChLP 4x2 1/2H

15.0203.203

	Name	Catalogue No.	Pcs
	Stand for instrument set for 3.0ChLP 4x2 1/2H	14.0203.203	1
	Guide rod 0.8/150	40.8132.150	4
	Guide rod 1.0/150	40.8133.150	4
	Guide rod 1.2/150	40.6466.150	4
	Dissecting forceps delicate 10.5cm	30.3376.000	1
	Elevator	40.4984.000	1
	Elevator	40.4986.000	1
	Hook	40.8477.000	1

Stand for plates 3.0ChLP 4x2 1/4H

14.0203.403

Plate thickness 1.2mm



Stand does not include implants



3.8010.006



3.8010.012



3.8011.005



3.8012.005



3.8013.005



3.8014.005



3.8015.007



3.8016.007



3.8017.007



3.8023.006

Stand for plates 3.0ChLP 4x2 1/4H

14.0203.404

Plate thickness 1.2mm



Stand does not include implants



3.8018.007



3.8019.004



3.8020.004



3.8021.003



3.8022.003



3.8024.002



3.8024.008



3.8025.004

Stand for plates 3.0ChLP 4x2 1/4H

14.0203.405

Plate thickness  0.8mm

Stand does not include implants



3.8010.206



3.8010.212



3.8011.205



3.8012.205



3.8013.205



3.8014.205



3.8015.207



3.8016.207



3.8017.207



3.8023.206

Stand for plates 3.0ChLP 4x2 1/4H

14.0203.406

Plate thickness  0.8mm

Stand does not include implants



3.8018.207



3.8019.204



3.8020.204



3.8021.203



3.8022.203



3.8024.202



3.8024.208



3.8025.204

STAND FOR PLATES - UNIVERSAL

Stand for plates 3,0ChLP 4x2 1/4H

(with cover)

14.0203.401




* Stand does not include implants

Stand for plates 3,0ChLP 4x2 1/4H

(without cover)


14.0203.402





* Stand does not include implants

Cover 4x2

14.0000.301




ARRANGEMENT EXAMPLE

OPTION I		OPTION II	
A set of stands for plates - one cover (<i>max 4</i>)		Separate stands for plates with covers	
			
Name	Catalogue No.	Name	Catalogue No.
1 Stand for plates 3,0ChLP 4x2 1/4H	14.0203.402	1 Stand for plates 3,0ChLP 4x2 1/4H	14.0203.401
2 Cover 4x2	14.0000.301		

9.c. STAND FOR SCREWS

Stand for 3.0ChLP screws 4x2H 1/2H

14.0203.501



Pcs




36

Stand does not include implants

Max screw length 20mm

ARRANGEMENT EXAMPLE

STANDS FOR 3.0ChLP IMPLANTS	Name	Catalogue No.
	1 Container lid 3.0ChLP 4x4H	14.0203.105
	2 Container for 3.0ChLP 4x4H	14.0203.104
	3 Stand for 3.0ChLP screws 4x2H 1/2H	14.0203.501
	4 Stand for plates 3.0ChLP 4x2 1/4H	14.0203.403
	5 Stand for plates 3.0ChLP 4x2 1/4H	14.0203.404
	6 Stand for plates 3.0ChLP 4x2 1/4H	14.0203.405
	7 Stand for plates 3.0ChLP 4x2 1/4H	14.0203.406



3.0ChLP straight plate




3.8010.120

20



3.0ChLP straight plate



3.8010.006


6

3.8010.012

12



3.0ChLP straight plate



3.8010.206

6

3.8010.212

12



3.0ChLP plate L



3.8011.005



3.8012.005



2/5



3.0ChLP plate L



3.8011.205



3.8012.205



2/5



3.0ChLP plate L 100°



3.8013.005



3.8014.005



2/5



3.0ChLP plate L 100°



3.8013.205



3.8014.205



2/5



3.0ChLP plate T



3.8016.007

2/7



3.0ChLP plate T




3.8016.207

2/7



3.0ChLP plate Y



3.8017.007

2/7



3.0ChLP plate Y



3.8017.207

2/7



3.0ChLP correction plate



3.8023.006

2/6



3.0ChLP correction plate




3.8023.206

2/6



3.0ChLP plate T



3.8015.007

3/7



3.0ChLP plate T



3.8015.207

3/7



3.0ChLP plate H




3.8018.007

3/7



3.0ChLP plate H



3.8018.207

3/7



3.0ChLP parallelogram plate



3.8019.004

2/4



3.8020.004



3.0ChLP parallelogram plate





3.8019.204


2/4





3.8020.204


Ti	3,0ChLP trapezium plate
1.2	
L	3.8021.003
R	3.8022.003
	2/3


Ti	3,0ChLP trapezium plate
0.8	
L	3.8021.203
R	3.8022.203
	2/3


Ti	3.0ChLP rectangular plate
1.2	
	3.8025.004
	2/4


Ti	3.0ChLP rectangular plate
0.8	
	3.8025.204
	2/4

Ti	3.0ChLP rectangular plate
1.2	
	3.8024.002 3.8024.008
	2/2 2/8


Ti	3.0ChLP rectangular plate
0.8	
	3.8024.202 3.8024.208
	2/2 2/8

TiA	Washer 1.2
1.2	
	3.1951.017
	5,5
	1

TiA	Washer 0.8
0.8	
	3.1951.016
	4,5
	1


Ti	Microplate
0,8	
	3.7309.000
Len	8
	1


3.0ChLP screw VA 1.5

Len	TiA	
6	3.8000.506	
7	3.8000.507	
8	3.8000.508	
9	3.8000.509	
10	3.8000.510	
11	3.8000.511	
12	3.8000.512	
13	3.8000.513	
14	3.8000.514	
15	3.8000.515	
16	3.8000.516	
17	3.8000.517	
18	3.8000.518	
19	3.8000.519	
20	3.8000.520	
Pcs	 5	




3.0ChLP screw VA 2.0


Len	TiA	
6	3.8001.506	
7	3.8001.507	
8	3.8001.508	
9	3.8001.509	
10	3.8001.510	
11	3.8001.511	
12	3.8001.512	
13	3.8001.513	
14	3.8001.514	
15	3.8001.515	
16	3.8001.516	
17	3.8001.517	
18	3.8001.518	
19	3.8001.519	
20	3.8001.520	
22*	3.8001.522	
24*	3.8001.524	
26*	3.8001.526	
28*	3.8001.528	
30*	3.8001.530	
Pcs	 5	



* Additionally Available


3.0ChLP screw VA 2.3

Len	TiA	
6	3.8002.506	
7	3.8002.507	
8	3.8002.508	
9	3.8002.509	
10	3.8002.510	
11	3.8002.511	
12	3.8002.512	
13	3.8002.513	
14	3.8002.514	
15	3.8002.515	
16	3.8002.516	
17	3.8002.517	
18	3.8002.518	
19	3.8002.519	
20	3.8002.520	
22*	3.8002.522	
24*	3.8002.524	
26*	3.8002.526	
28*	3.8002.528	
30*	3.8002.530	
Pcs	 5	




* Additionally Available


3.0ChLP screw 1.2


Len	TiA	
5*	3.8003.505	
6	3.8003.506	
7	3.8003.507	
8	3.8003.508	
9	3.8003.509	
10	3.8003.510	
11*	3.8003.511	
12*	3.8003.512	
13*	3.8003.513	
14*	3.8003.514	
15*	3.8003.515	
Pcs	 5	

* Additionally Available




3.0ChLP screw 1.5


Len	TiA	
6	3.8004.506	
7	3.8004.507	
8	3.8004.508	
9	3.8004.509	
10	3.8004.510	
11	3.8004.511	
12	3.8004.512	
13	3.8004.513	
14	3.8004.514	
15	3.8004.515	
16	3.8004.516	
17	3.8004.517	
18	3.8004.518	
19	3.8004.519	
20	3.8004.520	
Pcs	 5	




3.0ChLP screw 2.0

Len	TiA	
6	3.8005.506	
7	3.8005.507	
8	3.8005.508	
9	3.8005.509	
10	3.8005.510	
11	3.8005.511	
12	3.8005.512	
13	3.8005.513	
14	3.8005.514	
15	3.8005.515	
16	3.8005.516	
17	3.8005.517	
18	3.8005.518	
19	3.8005.519	
20	3.8005.520	
22*	3.8005.522	
24*	3.8005.524	
26*	3.8005.526	
28*	3.8005.528	
30*	3.8005.530	
Pcs	 5	

* Additionally Available



3.0ChLP screw 2.3

Len	TiA	
6	3.8006.506	
7	3.8006.507	
8	3.8006.508	
9	3.8006.509	
10	3.8006.510	
11	3.8006.511	
12	3.8006.512	
13	3.8006.513	
14	3.8006.514	
15	3.8006.515	
16	3.8006.516	
17	3.8006.517	
18	3.8006.518	
19	3.8006.519	
20	3.8006.520	
22*	3.8006.522	
24*	3.8006.524	
26*	3.8006.526	
28*	3.8006.528	
30*	3.8006.530	
Pcs	 5	

* Additionally Available



ChM sp. z o.o.

Lewickie 3b
16-061 Juchnowiec Kościelny
Poland
tel. +48 85 86 86 100
fax +48 85 86 86 101
chm@chm.eu
www.chm.eu



CE 0197